

H. pylori Ab SP Rapid Cassette Test Procedure How to Perform a Test - Cassette Test

Allow test cassette, specimen, and/or controls to equilibrate to room temperature (15-30°C) prior to testing.

1. Preparing Serum or Plasma Specimen:


- a. Separate serum or plasma from blood as soon as possible to avoid hemolysis. Use only clear non-hemolyzed specimens.
- **b.** Serum and plasma specimens may be stored at 2-8°C for up to 3 days. For long term storage, specimens should be kept below -20°C. Specimens should not be frozen and thawed repeatedly.

2. To Perform the Test:

- a. Remove the test cassette from the foil pouch and use it as soon as possible.
- b. Place the test cassette on a clean and level surface.
- c. Hold the dropper vertically and transfer 2 full drops of serum or plasma (approximately $80 \mu L$) to the Specimen Well of the test cassette and start the timer.
 - i. Avoid air bubbles.

3. Interpreting Results:


- a. Wait for the colored line(s) to appear.
- **b.** Read results at 10 minutes.
- c. Do not interpret the result after 20 minutes.





H. pylori Ab SP Rapid Cassette Test Procedure How to Perform a Test - Cassette Test

Interpretation of Results



- **POSITIVE**: Two distinct colored lines appear. One colored line should be in the control line region (C) and another apparent colored line should be in the test line region (T).
 - The intensity of the color in the test line region (T) will vary depending on the concentration of antibodies present in the specimen. Therefore, any shade of color in the test line region (T) should be considered positive.
- **NEGATIVE:** One colored line appears in the control line region (C). No line appears in the test line region (T).
- **INVALID:** Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test cassette. If the problem persists, discontinue using the test kit immediately and contact your local distributor.



ACON Laboratories, Inc. 10125 Mesa Rim Road, San Diego, CA 92121, U.S.A.

Tel: 1.858.875.8000 Fax: 1.858.200.0729 Email: info@aconlabs.com